

Datasheet: 4329-4906

BATCH NUMBER 151957

| | |
|----------------------|------------------------------|
| Description: | RABBIT ANTI ESCHERICHIA COLI |
| Specificity: | ESCHERICHIA COLI |
| Format: | Purified |
| Product Type: | Polyclonal Antibody |
| Isotype: | Polyclonal IgG |
| Quantity: | 1 ml |

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

| | Yes | No | Not Determined | Suggested Dilution |
|------------------|-----|----|----------------|--------------------|
| Flow Cytometry | | | ■ | |
| ELISA | ■ | | | 1/2000 |
| Western Blotting | | | ■ | |

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using the appropriate negative/positive controls.

| | |
|---------------------------------------|---|
| Target Species | Bacterial |
| Product Form | Purified IgG - liquid |
| Buffer Solution | Phosphate buffered saline |
| Preservative Stabilisers | 0.1% Sodium Azide (NaN ₃) |
| Approx. Protein Concentrations | IgG concentration 4.0 mg/ml |
| Immunogen | A mixture of all antigenic serotypes. |
| RRID | AB_619412 |
| Specificity | Rabbit anti <i>Escherichia coli</i> antibody recognizes <i>Escherichia coli</i> and is broadly |

reactive with all somatic and capsular (O and K) antigenic serotypes. The somatic O antigens are composed of lipopolysaccharide complexes which form part of the cell wall structure of *E. coli* whilst the capsular K antigens are mainly composed of acidic polysaccharide.

This antibody will remove *E. coli* proteins from recombinant preparations. Rabbit anti *Escherichia coli* antibody has not been absorbed and may cross-react with related enterobacteriaceae. Rabbit anti *Escherichia coli* antibody has been used in ELISA with serotypes O157:H7, O20, O125, O55, O111 and K12.

References

1. Tian, B. *et al.* (2015) Blu-ray optomagnetic measurement based competitive immunoassay for *Salmonella* detection. [Biosens Bioelectron. 77: 32-9.](#)
2. Skladal, P. *et al.* (2013) Electrochemical immunosensors for detection of microorganisms [Int. J. Electrochem. Sci., 8 1635 - 49](#)
3. Su, W.H. *et al.* (2013) Development of a chip-based multiplexed immunoassay using liposomal nanovesicles and its application in the detection of pathogens causing female lower genital tract infections. [Taiwan J Obstet Gynecol. 52: 25-32.](#)
4. Ho, T.Y. *et al.* (2013) Development of a novel bead-based 96-well filtration plate competitive immunoassay for the detection of Gentamycin. [Biosens Bioelectron. 49: 126-32.](#)
5. Ong, L.C. *et al.* (2014) Bacterial imaging with photostable upconversion fluorescent nanoparticles. [Biomaterials. 35 \(9\): 2987-98.](#)
6. Pivetal, J. *et al.* (2014) Micro-magnet arrays for specific single bacterial cell positioning [J. Magnetism and magnetic Materials 29 Sep \[Epub ahead of print\]](#)
7. Farka, Z. *et al.* (2015) Rapid Detection of Microorganisms Based on Active and Passive Modes of QCM [Sensors 15, 79-92](#)
8. Dayam RM *et al.* (2015) The Phosphoinositide-Gated Lysosomal Ca(2+) Channel, TRPML1, Is Required for Phagosome Maturation. [Traffic. 16 \(9\): 1010-26.](#)
9. Farka, Z. *et al.* (2015) Quartz crystal microbalance biosensor for rapid detection of aerosolized microorganisms. [Proc. SPIE 9455, Chemical, Biological, Radiological, Nuclear, and Explosives \(CBRNE\) Sensing XVI, 945507](#)
10. Bhokisham, N. *et al.* (2016) Modular Construction of Multi-Subunit Protein Complexes using Engineered Tags and Microbial Transglutaminase. [Metab Eng. May 26. pii: S1096-7176\(16\)30014-3. \[Epub ahead of print\]](#)
11. Kovář, D. *et al.* (2014) Detection of aerosolized biological agents using the piezoelectric immunosensor. [Anal Chem. 86 \(17\): 8680-6.](#)
12. VanGerven, N. *et al.* (2014) Secretion and functional display of fusion proteins through the curli biogenesis pathway. [Mol Microbiol. 91 \(5\): 1022-35.](#)
13. Rodrigues, D.M.C. *et al.* (2017) Sensitivity Analysis of Different Shapes of a Plastic Optical Fiber-Based Immunosensor for *Escherichia coli*: Simulation and Experimental Results. [Sensors \(Basel\). 17 \(12\) Dec 19 \[Epub ahead of print\].](#)

Storage

Store at +4°C or at -20°C if preferred.

Storage in frost-free freezers is not recommended.

This product should be stored undiluted. Avoid repeated freezing and thawing as this may denature the antibody. Should this product contain a precipitate we recommend microcentrifugation before use.

| | |
|--------------------------------------|--|
| Guarantee | 12 months from date of despatch |
| Health And Safety Information | Material Safety Datasheet documentation #10040 available at: https://www.bio-rad-antibodies.com/SDS/4329-4906 10040 |
| Regulatory | For research purposes only |

Related Products

Recommended Secondary Antibodies

Sheep Anti Rabbit IgG (STAR34...) [FITC](#)
Goat Anti Rabbit IgG (Fc) (STAR121...) [Biotin](#), [FITC](#), [HRP](#)
Sheep Anti Rabbit IgG (STAR35...) [RPE](#)
Goat Anti Rabbit IgG (H/L) (STAR124...) [HRP](#)

| | | | | | |
|----------------------------------|---|------------------|---|---------------|---|
| North & South America | Tel: +1 800 265 7376 Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com | Worldwide | Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com | Europe | Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com |
|----------------------------------|---|------------------|---|---------------|---|

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://www.bio-rad-antibodies.com/datasheets)
'M363084:200528'

Printed on 19 Jan 2024