

Datasheet: APO004

BATCH NUMBER 154355

Description:	pSIVA™ REAL-TIME APOPTOSIS FLUORESCENT MICROSCOPY KIT
Name:	pSIVA™ MICROSCOPY KIT
Other names:	ANNEXIN 12, ANNEXIN XII
Format:	IANBD (Green Fluorescence)
Product Type:	Kits
Quantity:	1 KIT

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Immunofluorescence	▪			Refer to Instructions For Use
Immunocytochemistry	▪			Refer to Instructions For Use
Live Cell Imaging	▪			Refer to Instructions For Use

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Product Information

The process of apoptosis is undertaken in several stages defined by specific cellular morphologies. One of the earlier stages of apoptosis is a change of the plasma membrane's phospholipid asymmetry. This rearrangement results in the translocation of phosphatidylserine (PS) from the inner to the outer plasma membrane (in non-apoptotic cells PS is exclusively located to the inner plasma membrane). However, apoptosis is reversible until reaching a certain point in the pathway and until then PS exposure can be considered as a transient event. The event defining whether the cell can be rescued and continues living is the onset of mitochondrial outer membrane permeabilization (Chipuk *et al.* 2006). Prior to reaching this point, PS exposure may be transient as molecules can relocate back to the inner plasma membrane (a phenomenon known as "PS flipping") ([van der Mark *et al.* 2013](#)).

The pSIVA™ (polarity-Sensitive Indicator of Viability & Apoptosis) probe is a biosensor conjugated to the green emitting IANBD dye (excitation maximum 488 nm, emission

maximum 530 nm) and only fluoresces when bound to PS in the presence of Ca²⁺ ([Kim et al. 2010a](#), [2010b](#)). The method thereby allows the analysis of kinetic apoptosis events in real time by live cell imaging and immunofluorescence / immunocytochemistry. In contrast to other PS detection based assays (e.g. annexin V) the pSIVA™ Real-Time Apoptosis Fluorescent Microscopy Kit does not require washing steps as you can simply add the probe and start analyzing.

Reagents In The Kit pSIVA-IANBD 200 µl
Propidium Iodide Staining Solution 500 µl

Instructions For Use Prior to commencing the microscopy experiment, please ensure that your cell culture medium contains between 1-2 mM Ca²⁺. Ca²⁺ is essential for binding of the pSIVA-IANBD probe to exposed phosphatidylserine ([Kim et al. 2010b](#)). If Ca²⁺ levels are insufficient, supplement the culture medium with 2 mM Ca²⁺.

1. Seed cells into culture plates and allow cells to adhere.
 2. Optional. After 24 hours exchange the culture medium for medium containing 2 mM Ca²⁺, if required.
 3. Optional. Induce apoptosis by treating cells with apoptosis inducing agents such as staurosporine or camptothecin.
 4. Add 10–20 µl/ml* of the pSIVA-IANBD probe to cells. Mix gently by moving culture plates backwards and forwards and side to side to ensure even distribution of the probe. **DO NOT PIPETTE TO MIX.**
 5. Optional. If distinction between apoptotic and necrotic/dead cells is desired, add between 5–10 µl/ml* of propidium iodide (PI) to cells. Mix gently by moving plates backwards and forwards and side to side to ensure even distribution of PI. **DO NOT PIPETTE TO MIX.**
 6. Observe cells under microscope using the green fluorescence filter for pSIVA-IANBD (excitation maximum 488 nm, emission maximum 530 nm) and the red fluorescence filter for PI (excitation maximum 535 nm, emission maximum 617 nm) visualization.
- * The stated pSIVA-IANBD and PI quantities are guidelines only and may have to be optimized.

Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/APO004.pdf

References

1. Kim, Y.E. *et al.* (2010) (a) Engineering a polarity-sensitive biosensor for time-lapse imaging of apoptotic processes and degeneration. [Nat Methods 7\(1\): 67–73.](#)
2. Kim, Y.E. *et al.* (2010) (b) Monitoring apoptosis and neuronal degeneration by real-time detection of phosphatidylserine externalization using a polarity-sensitive indicator of viability and apoptosis. [Nat Protoc. 5\(8\): 1396-405.](#)

Storage Store at +4°C. DO NOT FREEZE.
This product should be stored undiluted. This product is photosensitive and should be protected from light.

Guarantee 6 months from date of despatch

Acknowledgements pSIVA™ is a trademark of Novus Biologicals and is protected under patent no. 8.541.549.

Health And Safety Information Material Safety Datasheet documentation #10587 #10588 available at:
<https://www.bio-rad-antibodies.com/SDS/APO004>
Propidium Iodide Staining Solution (10587)
pSIVA-IANBD (10588)

Regulatory For research purposes only

Related Products

Recommended Useful Reagents

[ANNEXIN V:PE ASSAY KIT \(ANNEX50PE\)](#)

[ANNEXIN V:PE ASSAY KIT \(ANNEX200PE\)](#)

[ANNEXIN V:APC ASSAY KIT \(ANNEX50APC\)](#)

[ANNEXIN V:APC ASSAY KIT \(ANNEX200APC\)](#)

North & South Tel: +1 800 265 7376

America Fax: +1 919 878 3751

Email: antibody_sales_us@bio-rad.com

Worldwide

Tel: +44 (0)1865 852 700

Fax: +44 (0)1865 852 739

Email: antibody_sales_uk@bio-rad.com

Europe

Tel: +49 (0) 89 8090 95 21

Fax: +49 (0) 89 8090 95 50

Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://www.bio-rad-antibodies.com/datasheets)

'M349072:190228'

Printed on 12 Aug 2023

© 2023 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)