

# Datasheet: BUF026A

Description:	ANTIGEN RETRIEVAL BUFFER, pH10.0
Name:	ANTIGEN RETRIEVAL BUFFER, pH10.0 (BASIC)
Format:	Ready To Use
Product Type:	Accessory Reagent
Quantity:	500 ml

## **Product Details**

Applications	This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit <u>www.bio-rad-antibodies.com/protocols</u> .							
		Yes	No	Not Determined	Suggested Dilution			
	Immunohistology - Paraffin							
	Where this product has not been tested for use in a particular technique this does not							
	necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.							
Buffer Solution	Ready to use							
Intended Use	Bio-Rad's Antigen unmasking fluid is a heat induced, epitope retrieval system. The basic formulation is presented at pH10.0 and is suitable for use with antibodies that have not worked well with the universal formulation. Acidic and universal formulations are also available from Bio-Rad - see "recommended useful reagents" at the bottom of the datasheet for further information. The universal formulation is suitable for use with most antibodies and gives excellent preservation of tissue morphology. Like the basic form, the acidic form, may be of value for use with antibodies that do not work well with the universal formulation. These buffers may dramatically increase staining intensity but may also have an effect upon tissue morphology.							
Instructions For Use	1. Deparaffinize sections a	nd re-hyd	drate					
	2. Pre-heat retrieval buffer	to 92ºC	*.					
	3. Place slides in buffer and incubate 10 minutes at 92°C *							
	4. Allow to cool for 5-10 mi	inutes.						

	5. Rinse in distilled water or PBS, proceed with staining.				
	* N.B. Optimal unmasking is a function of time and temperature. Temperature (range 90°C-100°C) and time (5-30 minutes) should be determined by each investigator.				
Storage	Store at room temperature. Should this product contain a precipitate we recommend microcentrifugation before use.				
Guarantee	Guaranteed until date of expiry. Please see product label.				
Health And Safety Information	Material Safety Datasheet documentation #10093 available at: 10093: <u>https://www.bio-rad-antibodies.com/uploads/MSDS/10093.pdf</u>				
Regulatory	For research purposes only				

## Related Products

### **Recommended Useful Reagents**

#### ANTIGEN RETRIEVAL BUFFER, pH8.0 (BUF025A)

North & South	Tel: +1 800 265 7376	Worldwide	Tel: +44 (0)1865 852 700	Europe	Tel: +49 (0) 89 8090 95 21
America	Fax: +1 919 878 3751		Fax: +44 (0)1865 852 739		Fax: +49 (0) 89 8090 95 50
	Email: antibody_sales_us@bio-rad.com		Email: antibody_sales_uk@bio-rad.com		Email: antibody_sales_de@bio-rad.com

From March 15, 2021, we will no longer supply printed datasheets with our products. Look out for updates on how to access your digital version at bio-rad-antibodies.com 'M376144:210119'

### Printed on 09 Feb 2021

© 2021 Bio-Rad Laboratories Inc | Legal | Imprint