

# Datasheet: BUF033A BATCH NUMBER 168386

Description:	ELISA ULTRABLOCK	
Name:	ELISA ULTRABLOCK	
Format:	Ready To Use	
Product Type:	Accessory Reagent	
Quantity:	100 ml	

## **Product Details**

#### **Applications**

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit <a href="www.bio-rad-antibodies.com/protocols">www.bio-rad-antibodies.com/protocols</a>.

	Yes	No	Not Determined	Suggested Dilution
ELISA				Neat

Where this antibody has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the antibody for use in their own system using the appropriate negative/positive controls.

Product Form	Ready to use - liquid
Buffer Solution	Phosphate buffered saline
Preservative Stabilisers	0.05% Sodium Azide

#### **Intended Use**

BUF033A is a high performance ELISA blocking buffer for use in all ELISA formats with high background problems. The formulation includes a proprietary fish extract, molecular stabilizers and an antimicrobial agent. It is particularly useful in antigen-down and sandwich ELISA assays with mammalian samples, especially human, bovine and porcine serum, where the non-mammalian blocking proteins are less likely to interact with endogenous antibodies in the serum sample.

Superior blocking is possible due to the small size of the molecules which block non-specific binding sites on the adsorbed protein and unoccupied regions of the polystyrene plates often sterically inaccessible to traditional blockers. It provides a long-term stable environment for coating antigen or capture antibody. Plates can be blocked at room temperature and stored once dried for up to a year at +4°C.

#### **Instructions For Use**

- 1. Coat ELISA plate with antibody or antigen as required.
- 2. After incubation, remove the coating solution and wash the plate x2 with wash buffer. <u>BUF031A</u> can be used for this purpose.
- 3. Add 300-400ul of BUF033A and incubate for 2-24 hours. Use a volume equal to or greater than the volume of coating solution.
- 4. After removal of the blocking buffer continue with the assay or dry the plate for long-term storage at +4°C.

#### References

- 1. Colwell, D.D. *et al.* (2010) *Dicrocoelium dendriticum* in cattle from Cypress Hills, Canada: Humoral response and preliminary evaluation of an ELISA. <u>Vet Parasitol.174:</u> 162-165.
- 2. Defresne, F. *et al.* (2010) Differential influence of anticancer treatments and angiogenesis on the seric titer of autoantibody used as tumor and metastasis biomarker. Neoplasia 12: 562-570.
- 3. Zhang, Y. *et al.* (2012) Causes of alternative pathway dysregulation in dense deposit disease. Clin J Am Soc Nephrol. 7 (2): 265-74.
- 4. Krachudel, J. *et al.* (2013) Luteal insufficiency in bitches as a consequence of an autoimmune response against progesterone? Theriogenology. 79: 1278-83.
- 5. Yoshino, N. *et al.* (2013) Polymyxins as novel and safe mucosal adjuvants to induce humoral immune responses in mice. <u>PLoS One. 8 (4): e61643.</u>
- 6. Liao, S. *et al.* (2016) Human placental growth hormone is increased in maternal serum at 20 weeks of gestation in pregnancies with large-for-gestational-age babies. <u>Growth</u> Factors. 34 (5-6): 203-9.
- 7. Liao, S. *et al.* (2017) Maternal serum IGF-1, IGFBP-1 and 3, and placental growth hormone at 20weeks' gestation in pregnancies complicated by preeclampsia. <u>Pregnancy Hypertens</u>. 10: 149-54.
- 8. Vanhoutte, L. *et al.* (2017) MRI Assessment of Cardiomyopathy Induced by  $\beta$ 1-Adrenoreceptor Autoantibodies and Protection Through  $\beta$ 3-Adrenoreceptor Overexpression. <u>Sci Rep. 7: 43951.</u>
- 9. Ravindran, A. *et al.* (2018) C3 Glomerulopathy: Ten Years' Experience at Mayo Clinic. Mayo Clin Proc. 93 (8): 991-1008.
- 10. Zhang, Y. *et al.* (2020) Factor H Autoantibodies and Complement-Mediated Diseases. Front Immunol. 11: 607211.
- 11. Wang, Y. *et al.* (2021) Enhanced Bioactivity of a Human GHR Antagonist Generated by Solid-Phase Site-Specific PEGylation. Biomacromolecules. 22 (2): 299-308.

#### **Storage**

Store at +4°C.

DO NOT FREEZE

This product should be stored undiluted.

### Guarantee

Guaranteed until date of expiry. Please see product label.

#### **Health And Safety**

Material Safety Datasheet documentation #10379 available at:

Information <a href="https://www.bio-rad-antibodies.com/SDS/BUF033A">https://www.bio-rad-antibodies.com/SDS/BUF033A</a>

10379

**Regulatory** For research purposes only

## Related Products

# **Recommended Useful Reagents**

ELISA ULTRABLOCK (BUF033B)

ELISA ULTRABLOCK (BUF033C)

ELISA NEPTUNE ASSAY DILUENT (BUF039A)

5x ELISA COATING BUFFER (BUF030A)

10x ELISA WASH BUFFER (BUF031A)

 North & South
 Tel: +1 800 265 7376
 Worldwide
 Tel: +44 (0)1865 852 700
 Europe
 Tel: +49 (0) 89 8090 95 21

 America
 Fax: +1 919 878 3751
 Fax: +44 (0)1865 852 739
 Fax: +49 (0) 89 8090 95 50

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets 'M353954:190618'

#### Printed on 26 Jun 2024

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint