

Datasheet: BUF033C

BATCH NUMBER 167638

Description:	ELISA ULTRABLOCK
Name:	ELISA ULTRABLOCK
Format:	Ready To Use
Product Type:	Accessory Reagent
Quantity:	1000 ml

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
ELISA	▪			Neat

Where this antibody has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the antibody for use in their own system using the appropriate negative/positive controls.

Product Form	Ready to use - liquid
Buffer Solution	Phosphate buffered saline
Preservative Stabilisers	0.05% Sodium Azide

Intended Use BUF033C is a high performance ELISA blocking buffer for use in all ELISA formats with high background problems. The formulation includes a proprietary fish extract, molecular stabilizers and an antimicrobial agent. It is particularly useful in antigen-down and sandwich ELISA assays with mammalian samples, especially human, bovine and porcine serum, where the non-mammalian blocking proteins are less likely to interact with endogenous antibodies in the serum sample.

Superior blocking is possible due to the small size of the molecules which block non-specific binding sites on the adsorbed protein and unoccupied regions of the polystyrene plates often sterically inaccessible to traditional blockers. It provides a long-term stable environment for coating antigen or capture antibody. Plates can be blocked at room temperature and stored once dried for up to a year at +4°C.

- Instructions For Use**
1. Coat ELISA plate with antibody or antigen as required.
 2. After incubation, remove the coating solution and wash the plate x2 with wash buffer. [BUF031A](#) can be used for this purpose.
 3. Add 300-400ul of BUF033C and incubate for 2-24 hours. Use a volume equal to or greater than the volume of coating solution.
 4. After removal of the blocking buffer continue with the assay or dry the plate for long-term storage at +4°C.

- References**
1. Colwell, D.D. *et al.* (2010) *Dicrocoelium dendriticum* in cattle from Cypress Hills, Canada: Humoral response and preliminary evaluation of an ELISA. [Vet Parasitol.174: 162-165.](#)
 2. Defresne, F. *et al.* (2010) Differential influence of anticancer treatments and angiogenesis on the seric titer of autoantibody used as tumor and metastasis biomarker. [Neoplasia 12: 562-570.](#)
 3. Zhang, Y. *et al.* (2012) Causes of alternative pathway dysregulation in dense deposit disease. [Clin J Am Soc Nephrol. 7 \(2\): 265-74.](#)
 4. Krachudel, J. *et al.* (2013) Luteal insufficiency in bitches as a consequence of an autoimmune response against progesterone? [Theriogenology. 79: 1278-83.](#)
 5. Yoshino, N. *et al.* (2013) Polymyxins as novel and safe mucosal adjuvants to induce humoral immune responses in mice. [PLoS One. 8 \(4\): e61643.](#)
 6. Liao, S. *et al.* (2016) Human placental growth hormone is increased in maternal serum at 20 weeks of gestation in pregnancies with large-for-gestational-age babies. [Growth Factors. 34 \(5-6\): 203-9.](#)
 7. Liao, S. *et al.* (2017) Maternal serum IGF-1, IGFBP-1 and 3, and placental growth hormone at 20weeks' gestation in pregnancies complicated by preeclampsia. [Pregnancy Hypertens. 10: 149-54.](#)
 8. Vanhoutte, L. *et al.* (2017) MRI Assessment of Cardiomyopathy Induced by β 1-Adrenoreceptor Autoantibodies and Protection Through β 3-Adrenoreceptor Overexpression. [Sci Rep. 7: 43951.](#)
 9. Ravindran, A. *et al.* (2018) C3 Glomerulopathy: Ten Years' Experience at Mayo Clinic. [Mayo Clin Proc. 93 \(8\): 991-1008.](#)
 10. Zhang, Y. *et al.* (2020) Factor H Autoantibodies and Complement-Mediated Diseases. [Front Immunol. 11: 607211.](#)

Storage

Store at +4°C.

DO NOT FREEZE

This product should be stored undiluted.

Guarantee

Guaranteed until date of expiry. Please see product label.

Health And Safety Information

Material Safety Datasheet documentation #10379 available at: <https://www.bio-rad-antibodies.com/SDS/BUF033C>
10379

Related Products

Recommended Useful Reagents

[ELISA ULTRABLOCK \(BUF033A\)](#)

[ELISA ULTRABLOCK \(BUF033B\)](#)

[ELISA NEPTUNE ASSAY DILUENT \(BUF039A\)](#)

[5x ELISA COATING BUFFER \(BUF030A\)](#)

[10x ELISA WASH BUFFER \(BUF031A\)](#)

North & South Tel: +1 800 265 7376

America Fax: +1 919 878 3751

Email: antibody_sales_us@bio-rad.com

Worldwide

Tel: +44 (0)1865 852 700

Fax: +44 (0)1865 852 739

Email: antibody_sales_uk@bio-rad.com

Europe

Tel: +49 (0) 89 8090 95 21

Fax: +49 (0) 89 8090 95 50

Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets

'M353956:190618'

Printed on 23 Feb 2024