

Datasheet: BUF039A

BATCH NUMBER 155914

| | |
|----------------------|-----------------------------|
| Description: | ELISA NEPTUNE ASSAY DILUENT |
| Name: | ELISA NEPTUNE ASSAY DILUENT |
| Format: | Ready To Use |
| Product Type: | Accessory Reagent |
| Quantity: | 100 ml |

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

| | Yes | No | Not Determined | Suggested Dilution |
|-------|-----|----|----------------|--------------------|
| ELISA | ▪ | | | Neat |

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Product Form Ready to use - liquid

Buffer Solution Phosphate buffered saline

Preservative Stabilisers 0.1% Sodium Azide

Product Information **ELISA Neptune Assay Diluent** is an ELISA assay diluent designed to reduce background noise in problematic human, porcine and bovine serum or plasma samples. Specially formulated for use in antigen-down ELISA, it contains non-mammalian protein and proprietary detergent additives to reduce non-specific IgG adsorption to antigen coated and blocked ELISA well surfaces.

The difference in complexity, between serum or plasma samples and diluted standard curve proteins, is a potential cause of target analyte under-recovery. BUF039A equalizes the matrix, inhibits complement and thrombin activity and contains an antimicrobial agent.

Instructions For Use Pipette buffer directly onto the plate just prior to adding the sample or standard (suggested volume 50 - 100ul/well).

References

1. Defresne, F. *et al.* (2010) Differential influence of anticancer treatments and angiogenesis on the seric titer of autoantibody used as tumor and metastasis biomarker. [Neoplasia 12: 562-70.](#)
2. Massei, G. *et al.* (2015) Immunocontraception for managing feral cattle in Hong Kong. [PLoS One. 10 \(4\): e0121598.](#)

Storage

Store at +4°C.

DO NOT FREEZE

Guarantee

Guaranteed until date of expiry. Please see product label.

Health And Safety InformationMaterial Safety Datasheet documentation #10377 available at: <https://www.bio-rad-antibodies.com/SDS/BUF039A>
10377

Regulatory

For research purposes only

Related Products

Recommended Useful Reagents

[ELISA ULTRABLOCK \(BUF033A\)](#)[ELISA ULTRABLOCK \(BUF033B\)](#)[ELISA ULTRABLOCK \(BUF033C\)](#)[5x ELISA COATING BUFFER \(BUF030A\)](#)[10x ELISA WASH BUFFER \(BUF031A\)](#)[ELISA BSA BLOCK \(BUF032A\)](#)

North & South America Tel: +1 800 265 7376
Fax: +1 919 878 3751
Email: antibody_sales_us@bio-rad.com

Worldwide Tel: +44 (0)1865 852 700
Fax: +44 (0)1865 852 739
Email: antibody_sales_uk@bio-rad.com

Europe Tel: +49 (0) 89 8090 95 21
Fax: +49 (0) 89 8090 95 50
Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://www.bio-rad-antibodies.com/datasheets)
'M350297:190307'

Printed on 25 Mar 2023