Datasheet: BUF049C BATCH NUMBER 150786

HISPEC ASSAY DILUENT
HISPEC ASSAY DILUENT
Ready To Use
Accessory Reagent
N/A
500 ml

Product Details

Applications	This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit <u>www.bio-rad-antibodies.com/protocols</u> .								
		Yes	No	Not Determined	Suggested Dilution				
	ELISA	•							
	Immunoassay	•							
	Where this product has not been tested for use in a particular technique this does no necessarily exclude its use in such procedures. Suggested working dilutions are give a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.								
Product Form	Ready to use assay diluent buffer								
Preservative Stabilisers	0.1% Proclin™ 300								
Intended Use	HISPEC assay diluent works to reduce cross reactivity, non-specific binding and matrix effects in immunoassays such as ELISA, EIA, Western blotting, immuno-PCR, protein arrays, multianalyte immunoassays and immunohistochemistry.								
	The effect of HISPEC assay diluent is dependent on assay system and the antibodies used. HISPEC assay diluent is used instead of a sample buffer or antibody dilution buffer within the immunoassay protocol.								
	king of surfaces. For bl ithin the Bio-Rad range	•							

ELISA: use as a dilution buffer for the specimen and for the detection antibodies Western blotting: use as dilution buffer for primary and secondary antibodies Immunohistochemistry: use as dilution buffer for primary and secondary antibodies
Protein arrays: use as dilution buffer for specimen and for the detection antibodies
Instructions For Use The buffer should be mixed thoroughly immediately before use. Dilution of the samples: The standards and samples for ELISA and protein arrays can be diluted with HISPEC assay diluent at 1:2 or higher. Standards and samples should be treated identically.
Dilution of antibodies: Antibodies can be diluted with HISPEC assay diluent in a user-defined manner, depending on the recommendation of the antibody supplier.
Expected results: In some cases a reduction in assay signal intensity may be observed when using HISPEC assay diluent. This does not result in a reduction in assay sensitivity due to the accompanying reduction in non-specific binding.
Signal intensity may be increased in some assay systems by use of higher concentrations of antibodies, or by dilution of the HISPEC assay diluent in distilled water (1:2) prior to use. However, these adjustments may also result in some associated increase in non-specific binding.
References 1. Eriksson, O. et al. (2017) Pancreatic imaging using an antibody fragment targeting the zinc transporter type 8: a direct comparison with radio-iodinated Exendin-4. Acta Diabetol. Oct 24 [Epub ahead of print].
StorageStore at +4°C or at -20°C if preferred.Storage in frost-free freezers is not recommended.This product should be stored undiluted. Avoid repeated freezing and thawing.
Guarantee Guaranteed until date of expiry. Please see product label.
Acknowledgements Proclin [™] 300 is a trademark of The Dow Chemical Company ("Dow") or an affiliated company of Dow.
Health And Safety Material Safety Datasheet documentation #10265 available at: Information <u>https://www.bio-rad-antibodies.com/SDS/BUF049C</u> 10265
Regulatory For research purposes only

Related Products

Recommended Useful Reagents

North & South	Tel: +1 800 265 7376	Worldwide	Tel: +44 (0)1865 852 700	Europe	Tel: +49 (0) 89 8090 95 21
America	Fax: +1 919 878 3751		Fax: +44 (0)1865 852 739		Fax: +49 (0) 89 8090 95 50
	Email: antibody_sales_us@)bio-rad.com	Email: antibody_sales_uk@bio	o-rad.com	Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets

'M349767:190307'

Printed on 29 Feb 2024

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint