

Datasheet: DC012

BATCH NUMBER 153582

Description:	MOUSE IgG1:FITC/MOUSE IgG1:RPE NEGATIVE CONTROL
Specificity:	MOUSE IgG1/IgG1 NEGATIVE CONTROL
Format:	FITC/RPE
Product Type:	Negative/Isotype Control
Isotype:	Cocktail
Quantity:	50 TESTS

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	▪			Neat

Where this antibody has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the antibody for use in their own system using appropriate negative/positive controls.

Antibody Isotypes

FITC reagent: IgG1 (MOUSE)
RPE reagent: IgG1 (MOUSE)

Target Species

Negative Control

Product Form

Dual Colour combination consisting of FITC conjugated and RPE conjugated monoclonal antibodies mixed in optimal ratio - lyophilised.

Reconstitution

Reconstitute with 0.5 ml distilled water

Max Ex/Em	Fluorophore	Excitation Max (nm)	Emission Max (nm)
	FITC	490	525
	RPE 488nm laser	496	578
	RPE 561nm laser	546	578

Buffer Solution

Phosphate buffered saline

Preservative Stabilisers

0.09% Sodium Azide
1% Bovine Serum Albumin

5% Sucrose

RRID	AB_322251
-------------	-----------

Specificity	Mouse IgG1 FITC / Mouse IgG1 RPE negative control is suitable for use as a negative control for the measurement of non-specific binding of mouse monoclonals of isotype IgG1 to human tissues in a dual labelling technique using FITC and R-Phycoerythrin fluorochromes.
--------------------	--

Flow Cytometry	Use 10ul of the suggested working dilution to label 10 ⁶ cells or 100ul whole blood.
-----------------------	---

References	<ol style="list-style-type: none">1. Steele, J. <i>et al.</i> (2002) Detection of CD4(+)- and CD8(+)- T-cell responses to human papillomavirus type 1 antigens expressed at various stages of the virus life cycle by using an enzyme-linked immunospot assay of gamma Interferon release. J. Virol. 76: 6027 - 6036.2. Youn, S.W. <i>et al.</i> (2004) Cellular senescence induced loss of stem cell proportion in the skin in vitro. J Dermatol Sci. 35: 113-23.
-------------------	---

Storage	Prior to reconstitution store at +4°C. Following reconstitution store at +4°C This product should be stored undiluted. DO NOT FREEZE. This product is photosensitive and should be protected from light. Should this product contain a precipitate we recommend microcentrifugation before use.
----------------	---

Guarantee	12 months from date of despatch
------------------	---------------------------------

Health And Safety Information	Material Safety Datasheet documentation #20487 available at: https://www.bio-rad-antibodies.com/SDS/DC012 20487
--------------------------------------	---

Regulatory	For research purposes only
-------------------	----------------------------

North & South America	Tel: +1 800 265 7376 Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com	Worldwide	Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com	Europe	Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com
----------------------------------	---	------------------	---	---------------	---

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://www.bio-rad-antibodies.com/datasheets)
'M375208:210104'

Printed on 12 Aug 2023