

Datasheet: FCSC555A

BATCH NUMBER 160628

Description:	QUANTUM™ FITC-5 MESF
Name:	QUANTUM™ FITC-5 MESF
Format:	Flow Cytometry Calibration Reagent
Product Type:	Accessory Reagent
Quantity:	20 TESTS

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	▪			

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Buffer Solution	Phosphate buffered saline
------------------------	---------------------------

Preservative Stabilisers	0.05% Sodium Azide (NaN ₃) ≤0.05% Tween 80
---------------------------------	---

Product Information	Quantum™ FITC-5 MESF is comprised of five calibrated fluorescent populations, with different levels of FITC fluorescence intensity, and one blank population of uniform microspheres that are approximately the size of human lymphocytes (7-9µm). The excitation and emission spectra of the microspheres match those of cell samples labeled with FITC.
----------------------------	--

Intended Use	FCSC555A is used in the quantitation of FITC fluorescence intensity in Molecules of Equivalent Soluble Fluorochrome (MESF) units. When used in conjunction with Simply Cellular® microspheres this kit also allows quantitation of Antibody Binding Capacity (ABC).
---------------------	---

This kit enables fluorescence intensity of a sample to be directly quantified in terms of MESF units and for samples from different instruments to be accurately compared. The fluorescence intensity of each of the populations in the kit has been calibrated against solutions of laser grade fluorescent dye in units of MESF FITC per microsphere. The

Certified Blank™ population is used to measure the fluorescence detection threshold of the instrument. Correct use of the kit enables:

- 1) quantitation of fluorescence intensity of samples in terms of MESF;
- 2) determination of instrument fluorescence detection threshold;
- 3) determination of instrument linearity;
- 4) data comparison over time and between multiple instruments.

Reagents In The Kit 1 x 1ml bottle of unlabeled microbeads
5 x 1ml bottles of labeled microbeads

Instructions For Use Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/FCSC555A.pdf

Storage Store at +4°C. DO NOT FREEZE.
This product should be stored undiluted. This product is photosensitive and should be protected from light.

Guarantee Guaranteed until date of expiry. Please see product label.

Acknowledgements Quantum™ is a trademark of Bangs Laboratories, INC.

Health And Safety Information Material Safety Datasheet documentation #10042 available at:
10042: <https://www.bio-rad-antibodies.com/uploads/MSDS/10042.pdf>

Regulatory For research purposes only

North & South Tel: +1 800 265 7376

America Fax: +1 919 878 3751

Email: antibody_sales_us@bio-rad.com

Worldwide

Tel: +44 (0)1865 852 700

Fax: +44 (0)1865 852 739

Email: antibody_sales_uk@bio-rad.com

Europe

Tel: +49 (0) 89 8090 95 21

Fax: +49 (0) 89 8090 95 50

Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets
'M350333:190307'

Printed on 07 Feb 2023

© 2023 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)