

Datasheet: FCSC580

BATCH NUMBER 170371

Description:	FLOW CYTOMETRY ABSOLUTE COUNT STANDARD™
Name:	FLOW CYTOMETRY ABSOLUTE COUNT STANDARD™
Format:	Flow Cytometry Validation Reagent
Product Type:	Accessory Reagent
Quantity:	20 TESTS

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	▪			

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Buffer Solution Phosphate buffered saline.

Preservative 0.01% Gelatin
Stabilisers ≤ 0.05% Tween 20
 0.09% Sodium Azide (NaN₃)

Product Information **Flow Cytometry Absolute Count Standard** is a suspension of microspheres that are approximately the size of human lymphocytes (7-9µm) and are presented as a known concentration. They are internally labeled with multiple fluorochromes and the combination of dyes allows the beads to be excited by a common argon laser (488nm) and emit in the three standard channels of a flow cytometer (FL1, FL2, FL3).

Intended Use **Flow Cytometry Absolute Count Standard** is intended for use as an internal counting standard. It is designed for use in the proper set-up of flow cytometers and cell counters and for the accurate enumeration of cells or particles.

Reagents In The Kit 1 bottle containing 10ml of microspheres.

Instructions For Use Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/FCSC580.pdf

References 1. McDonald, J.U. *et al.* (2011) *In vivo* functional analysis and genetic modification of *in vitro*-derived mouse neutrophils. [FASEB J. 25 \(6\): 1972-82.](#)
2. Dunsterville, C. *et al.* (2019) The Use of Dual-Cell-Tracker Dye Staining for the Identification and Characterization of Peanut-Specific T-Cell Subsets. [Methods Mol Biol. 2020: 143-52.](#)

Storage Store at +4°C. DO NOT FREEZE.
This product should be stored undiluted. This product is photosensitive and should be protected from light.

Guarantee Guaranteed until date of expiry. Please see product label.

Acknowledgements Absolute Count Standard is a trademark of Bangs Laboratories, INC.

Health And Safety Information Material Safety Datasheet documentation #10042 available at:
<https://www.bio-rad-antibodies.com/SDS/FCSC580>
10042

Regulatory For research purposes only.

North & South America	Tel: +1 800 265 7376 Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com	Worldwide	Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com	Europe	Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com
----------------------------------	---	------------------	---	---------------	---

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://www.bio-rad-antibodies.com/datasheets)
'M420801:230706'

Printed on 23 May 2025

© 2025 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)