

Datasheet: ICT9158 BATCH NUMBER 165135

Description:	PYROPTOSIS 660 CASPASE-1 KIT
Name:	PYROPTOSIS 660 CASPASE-1
Format:	660 (Red Fluorescence)
Product Type:	Kits
Quantity:	25 TESTS

Product Details

Applications	This product has been reported to work in the following applications. This information is
	derived from testing within our laboratories, peer-reviewed publications or personal
	communications from the originators. Please refer to references indicated for further
	information. For general protocol recommendations, please visit <u>www.bio-</u>
	rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	-			Refer to Instructions For Use
Immunofluorescence	•			

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Max Ex/Em	Fluorophore	Excitation Max (nm)	Emission Max (nm)
	FLICA 660	660	685

Product InformationPyroptosis 660 Caspase-1 Kit utilizes the popular FLICA technology to detect caspase-1
activation. This kits contain the caspase-1 inhibitor reagent YVAD-FMK, which has the
preferred binding sequence for caspase-1, Tyr-Val-Ala-Asp (YVAD) (Chapman, 1992).
This preferred caspase-1 binding sequence is labeled with 660 a far red fluorescent dye
and linked to a fluoromethyl ketone (FMK) reactive entity. Caspase-1 will not cleave the
FLICA inhibitor probe; instead, it forms an irreversible covalent bond with the FMK group
on the reagent and becomes inhibited from further enzymatic activity.

Test PrincipleTo use FLICA, add directly to the cell culture medium, incubate, and wash. FLICA is
cell-permeant and will efficiently diffuse in and out of all cells. If there is an active
caspase-1 enzyme inside the cell, it will covalently bind with YVAD-FMK and retain the
fluorescent signal within the cell. Unbound FLICA will diffuse out of the cell during the
subsequent wash steps. Therefore, positive cells will retain a higher concentration of
FLICA and fluoresce brighter than negative cells. There is no interference from

		counter-stained with c be counter-stained wi 7-AAD. Nuclear morpl	g with FLICA, cells can be s labeled with YVAD-FMK can stains Propidium lodide and using Hoechst 33342 iewed through a fluorescence			
Reagents	s In The Kit	1 vial of 660-YVAD-FI				
		1 vial Nigericin 10x Cellular Wash Bu Fixative, 6 mL 1 vial Hoechst Stain, 7				
Instructio	ons For Use	Instructions for use can be found at https://www.bio-rad-antibodies.com/static/uploads/lifu/ict9158.pdf				
Storage		each unopened comp label. Store the Nigeri	E CONDITIONS APPLY ON ARRIVAL. onent) according to the storage instruc cin at -20°C. Once reconstituted, the N ted and stored at -20°C for 12 months	tions on each component Nigericin stock should be used		
Guarante	e	Guaranteed until date	of expiry. Please see product label.			
Acknowle	edgements	FLICA is a trademark of Immunochemistry Technologies, LLC.				
	Health And Safety InformationMaterial Safety Datasheet documentation #20374 #10476 #2 available at: https://www.bio-rad-antibodies.com/SDS/ICT9158 660-YVAD-FMK caspase-1 inhibitor reagent (20374) Hoechst Stain (10476) Nigericin (20431) 10X Cellular Wash Buffer (20435) Fixative (10498)		20431 #20435 #10498			
Regulato	Regulatory For research purposes only					
North & South America	orth & South Tel: +1 800 265 7376 Worldwide nerica Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com		Tel: +44 (0)1865 852 700 Europe Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com	Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com		
To find a b	atch/lot speci	fic datasheet for this produ	uct, please use our online search tool at: b 'M417811:230411'	io-rad-antibodies.com/datasheets		
			Printed on 17 Apr 2024			

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint