

Datasheet: ICT943 BATCH NUMBER 165042

MITOCHONDRIAL PERMEABILITY TRANSITION KIT
MITOCHONDRIAL PERMEABILITY TRANSITION: MitoPT JC-1 KIT
JC-1 Dye (Dual Green/Red Fluorescence)
Kits
100 TESTS

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	•			
Immunofluorescence				

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Product Information

The MitoPT JC-1 kit uses a quick and easy staining method to clearly differentiate between non-apoptotic and apoptotic cells through mitochondrial functionality.

Test Principle

An early indication of apoptosis involves a collapse in the electrochemical gradient across the mitochondrial membrane. Loss of mitochondrial membrane potential can be detected by a unique fluorescent cationic dye known as JC-1 (5,5',6,6'-tetrachloro-1,1',3,3'-tetraethylbenzamidazolocarbocyanin iodide) that has been incorporated into the MitoPT JC-1 kit.

The MitoPT JC-1 reagent easily penetrates cells and enters the mitochondria. It aggregates in the mitochondria of non-apoptotic cells and fluoresces red, whilst in apoptotic cells it diffuses throughout the cell. Once dispersed, the reagent assumes a monomeric form and fluoresces green. This allows an easy distinction between non-apoptotic red fluorescent cells and apoptotic green fluorescent cells which can be read with a flow cytometer, fluorescence microscope or a fluorescence plate reader using black microtitre plates.

Reagents In The Kit

MitoPT JC-1 Reagent, 100 tests 10x Assay Buffer, 60 mL CCCP (50 mM in DMSO), 125 ul.

Instructions For Use	Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/ICT943.pdf
References	1. Jasek, E. et al. (2012) Effect of histone deacetylase inhibitors trichostatin A and valproid
	acid on etoposide-induced apoptosis in leukemia cells. Anticancer Res. 32 (7): 2791-9.
	2. Wei, X. et al. (2014) In Vitro Comparative Effect of Three Novel Borate Bioglasses on
	the Behaviors of Osteoblastic MC3T3-E1 Cells J Materials Sci Tech 30 (10): 979-83.
	3. Gizak, A. <i>et al.</i> (2019) Targeting a moonlighting function of aldolase induces apoptosis
	in cancer cells. Cell Death Dis. 10 (10): 712.
Storage	Store the unopened kit and each unopened component at -20°C until the expiration date.
	Once opened, some components may be stored at 2-8°C until the expiration date. CCCP
	should be stored frozen. Once reconstituted with DMSO, dilute and use MitoPT reagent
	immediately, or store at ≤-20°C for 12 months protected from light and thawed no more
	than twice.
Guarantee	Guaranteed until date of expiry. Please see product label.
Acknowledgements	MitoPT is a trademark of Immunochemistry Technologies, LLC.
Health And Safety	Material Safety Datasheet documentation #20291 #10474 #10479 available at:
Information	https://www.bio-rad-antibodies.com/SDS/ICT943
	MitoPT JC-1 Reagent (20291)
	10x Assay Buffer (10474)
	CCCP Reagent (10479)
Regulatory	For research purposes only.

Related Products

Recommended Useful Reagents

MITOCHONDRIAL PERMEABILITY TRANSITION KIT (ICT944)

North & South Tel: +1 800 265 7376

America Fax: +1 919 878 3751 Worldwide

Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Europe

Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50

Email: antibody_sales_us@bio-rad.com

Email: antibody_sales_uk@bio-rad.com

Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets 'M404840:220906'

Printed on 19 Jan 2024

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint