

## Datasheet: MCA1209F

**BATCH NUMBER 165681**

|                      |                                  |
|----------------------|----------------------------------|
| <b>Description:</b>  | MOUSE IgG1 NEGATIVE CONTROL:FITC |
| <b>Specificity:</b>  | MOUSE IgG1 NEGATIVE CONTROL      |
| <b>Format:</b>       | FITC                             |
| <b>Product Type:</b> | Negative/Isotype Control         |
| <b>Isotype:</b>      | IgG1                             |
| <b>Quantity:</b>     | 0.1 mg                           |

## Product Details

### Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit [www.bio-rad-antibodies.com/protocols](http://www.bio-rad-antibodies.com/protocols).

|                | Yes | No | Not Determined | Suggested Dilution |
|----------------|-----|----|----------------|--------------------|
| Flow Cytometry | ■   |    |                | *                  |

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

|                                       |   |                            |                          |
|---------------------------------------|---|----------------------------|--------------------------|
| <b>Target Species</b>                 | Negative Control  |                            |                          |
| <b>Product Form</b>                   | Purified IgG conjugated to Fluorescein Isothiocyanate Isomer 1 (FITC) - liquid                |                            |                          |
| <b>Max Ex/Em</b>                      | <b>Fluorophore</b>  | <b>Excitation Max (nm)</b> | <b>Emission Max (nm)</b> |
|                                       | FITC  | 490                        | 525                      |
| <b>Preparation</b>                    | Purified IgG prepared by affinity chromatography on Protein A from tissue culture supernatant |                            |                          |
| <b>Buffer Solution</b>                | Phosphate buffered saline   |                            |                          |
| <b>Preservative Stabilisers</b>       | 0.09% sodium azide (NaN <sub>3</sub> )  |                            |                          |
|                                       | 1% bovine serum albumin   |                            |                          |
| <b>Approx. Protein Concentrations</b> | IgG concentration 0.1 mg/ml   |                            |                          |
| <b>Immunogen</b>                      | Human T lymphocytes.  |                            |                          |

|                                      |  |
|--------------------------------------|--|
| <b>RRID</b>                          | AB_567337  |
| <b>Fusion Partners</b>               | Spleen cells from immunized BALB/c mice were fused with cells of the NS1 mouse myeloma cell line.  |
| <b>Specificity</b>                   | <p><b>Mouse IgG1 Negative Control antibody</b> is suitable for use as a negative control to assess non-specific binding of mouse IgG1 antibodies to target cells. Mouse IgG1 Negative Control antibody has been tested and found to be negative on the following rat cell types, peripheral blood leucocytes, thymocytes, splenocytes and macrophages.</p> <p>Clone F8-11-13 recognizes the human CD45RA antigen, and therefore human leucocytes may be used as a positive control for this product. NOT SUITABLE FOR USE AS A NEGATIVE CONTROL ON HUMAN TISSUES</p>   |
| <b>Flow Cytometry</b>                | Use 10µl of the suggested working dilution to label 10 <sup>6</sup> cells in 100µl   |
| <b>References</b>                    | <ol style="list-style-type: none"> <li>1. Weiss, D.J. <i>et al.</i> (2008) Bovine monocyte TLR2 receptors differentially regulate the intracellular fate of <i>Mycobacterium avium</i> subsp. <i>paratuberculosis</i> and <i>Mycobacterium avium</i> subsp. <i>avium</i>. <a href="#">J Leukoc Biol. 83 (1): 48-55.</a></li> <li>2. Chen, W. <i>et al.</i> (2009) Expression of toll-like receptor 4 in uvea-resident tissue macrophages during endotoxin-induced uveitis. <a href="#">Mol Vis. 15: 619-28.</a></li> <li>3. Safeukui I <i>et al.</i> (2015) Malaria induces anemia through CD8+ T cell-dependent parasite clearance and erythrocyte removal in the spleen. <a href="#">MBio. 6 (1) pii: e02493-14.</a></li> <li>4. Aricha, R. <i>et al.</i> (2016) Suppression of experimental autoimmune myasthenia gravis by autologous T regulatory cells. <a href="#">J Autoimmun. 67: 57-64.</a></li> <li>5. Wattedgedera, S.R. <i>et al.</i> (2017) Enhancing the toolbox to study IL-17A in cattle and sheep. <a href="#">Vet Res. 48 (1): 20.</a></li> <li>6. Stangl, H. <i>et al.</i> (2020) MHC/class-II-positive cells inhibit corticosterone of adrenal gland cells in experimental arthritis: a role for IL-1β, IL-18, and the inflammasome. <a href="#">Sci Rep. 10 (1): 17071.</a></li> <li>7. Terpeluk, R.E. <i>et al.</i> (2024) Supplementation of Foals with a <i>Saccharomyces cerevisiae</i> Fermentation Product Alters the Early Response to Vaccination <a href="#">Animals. 14 (6): 960.</a></li> </ol> |
| <b>Storage</b>                       | <p>This product is shipped at ambient temperature. It is recommended to aliquot and store at -20°C on receipt. When thawed, aliquot the sample as needed. Keep aliquots at 2-8°C for short term use (up to 4 weeks) and store the remaining aliquots at -20°C.</p> <p>Avoid repeated freezing and thawing as this may denature the antibody. Storage in frost-free freezers is not recommended. This product is photosensitive and should be protected from light.</p>   |
| <b>Guarantee</b>                     | 12 months from date of despatch  |
| <b>Health And Safety Information</b> | <p>Material Safety Datasheet documentation #10041 available at: <a href="https://www.bio-rad-antibodies.com/SDS/MCA1209F10041">https://www.bio-rad-antibodies.com/SDS/MCA1209F10041</a></p>  |
| <b>Regulatory</b>                    | For research purposes only   |

---

## Related Products

### Recommended Negative Controls

[MOUSE IgG1 NEGATIVE CONTROL:FITC \(MCA928F\)](#)

**North & South** Tel: +1 800 265 7376

**America** Fax: +1 919 878 3751

Email: [antibody\\_sales\\_us@bio-rad.com](mailto:antibody_sales_us@bio-rad.com)

**Worldwide**

Tel: +44 (0)1865 852 700

Fax: +44 (0)1865 852 739

Email: [antibody\\_sales\\_uk@bio-rad.com](mailto:antibody_sales_uk@bio-rad.com)

**Europe**

Tel: +49 (0) 89 8090 95 21

Fax: +49 (0) 89 8090 95 50

Email: [antibody\\_sales\\_de@bio-rad.com](mailto:antibody_sales_de@bio-rad.com)

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://bio-rad-antibodies.com/datasheets)

'M408486:221012'

**Printed on 25 Mar 2024**

---

© 2024 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)