

## Datasheet: MCA929B

<b>Description:</b>	MOUSE IgG2a NEGATIVE CONTROL:Biotin
<b>Specificity:</b>	MOUSE IgG2a NEGATIVE CONTROL
<b>Format:</b>	Biotin
<b>Product Type:</b>	Negative/Isotype Control
<b>Isotype:</b>	IgG2a
<b>Quantity:</b>	100 TESTS

## Product Details

### Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit [www.bio-rad-antibodies.com/protocols](http://www.bio-rad-antibodies.com/protocols).

	Yes	No	Not Determined	Suggested Dilution
Flow Cytometry	▪			*
Immunohistology - Frozen			▪	
Immunohistology - Paraffin			▪	
ELISA			▪	
Western Blotting			▪	

Where this antibody has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures.

\*It is recommended that the user dilutes the antibody for use in their own system to a concentration equivalent to their test reagents.

<b>Target Species</b>	Negative Control
<b>Product Form</b>	Purified IgG conjugated to Biotin - liquid
<b>Preparation</b>	Purified IgG prepared by affinity chromatography from tissue culture supernatant
<b>Buffer Solution</b>	Phosphate buffered saline
<b>Preservative Stabilisers</b>	0.09% Sodium Azide 1% Bovine Serum Albumin
<b>Approx. Protein Concentrations</b>	IgG concentration 0.1 mg/ml
<b>Immunogen</b>	Activated rat T-helper cells.

RRID AB\_322270

---

**Fusion Partners** Spleen cells from immunised BALB/c mice were fused with cells of the NS1 mouse myeloma cell line.

---

**Specificity** **Mouse IgG2a negative control antibody, clone OX-34** is suitable for use as a negative control reagent for the measurement of non-specific binding of mouse monoclonal antibodies of isotype IgG2a to human tissue.

Clone MRC OX-34 recognises a rat cell surface marker, and therefore cannot be used as a negative control in this species.

This product is routinely tested in flow cytometry on rat splenocytes to confirm antibody activity and on human whole blood to test for suitability as a negative control.

Test results have shown that MCA929 is also suitable for use as a negative control with bovine, ovine, porcine, equine, canine, lapine and guinea-pig tissues.

This antibody may not be suitable for intracellular staining on some cell types.

---

**Flow Cytometry** Use 10ul of the suggested working dilution to label  $10^6$  in 100ul.

---

**References**

1. Avigdor, A. *et al.* (2004) CD44 and hyaluronic acid cooperate with SDF-1 in the trafficking of human CD34+ stem/progenitor cells to bone marrow. [Blood. 103 \(8\): 2981-9.](#)
2. Kamble, N.M. *et al.* (2016) Interaction of a live attenuated *Salmonella gallinarum* vaccine candidate with chicken bone marrow-derived dendritic cells. [Avian Pathol. Jan 26:1-24. \[Epub ahead of print\]](#)
3. Wattedgedera, S.R. *et al.* (2017) Enhancing the toolbox to study IL-17A in cattle and sheep. [Vet Res. 48 \(1\): 20.](#)

---

**Storage** This product is shipped at ambient temperature. It is recommended to aliquot and store at -20°C on receipt. When thawed, aliquot the sample as needed. Keep aliquots at 2-8°C for short term use (up to 4 weeks) and store the remaining aliquots at -20°C.

Avoid repeated freezing and thawing as this may denature the antibody. Storage in frost-free freezers is not recommended.

---

**Guarantee** 12 months from date of despatch

---

**Health And Safety Information** Material Safety Datasheet documentation #10041 available at: <https://www.bio-rad-antibodies.com/SDS/MCA929B>  
10041

---

**Regulatory** For research purposes only

---

**North & South America** Tel: +1 800 265 7376  
Fax: +1 919 878 3751

Email: [antibody\\_sales\\_us@bio-rad.com](mailto:antibody_sales_us@bio-rad.com)

**Worldwide** Tel: +44 (0)1865 852 700  
Fax: +44 (0)1865 852 739

Email: [antibody\\_sales\\_uk@bio-rad.com](mailto:antibody_sales_uk@bio-rad.com)

**Europe** Tel: +49 (0) 89 8090 95 21  
Fax: +49 (0) 89 8090 95 50

Email: [antibody\\_sales\\_de@bio-rad.com](mailto:antibody_sales_de@bio-rad.com)

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://bio-rad-antibodies.com/datasheets)

'M410391:221028'

**Printed on 25 Mar 2023**

---

© 2023 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)