

Datasheet: OBT1999G

BATCH NUMBER 154508

Description:	ELISA NEPTUNE SAMPLE DILUENT
Name:	ELISA NEPTUNE SAMPLE DILUENT
Format:	Ready To Use
Product Type:	Accessory Reagent
Quantity:	100 ml

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
ELISA	▪			Neat

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Product Form Ready to use - liquid.

Preservative Stabilisers 0.1% Sodium Azide (NaN₃)

Intended Use **ELISA Neptune Sample Diluent** is a non-mammalian protein based diluent highly recommended for use with serum or plasma samples from mouse, porcine or bovine sources in antigen-down ELISAs, and human or other mammalian samples in sandwich ELISAs. OBT1999G dilutes highly concentrated samples into the functional range of the assay, preventing overloading of the coating proteins. It also performs well as a protein diluent in standard curve preparation. The proprietary additives in a buffered PBS base minimize non-specific adsorption of serum IgG to the coated plate. This can be an issue with the highly glycosylated “sticky” porcine serum IgG, leading to high background noise. An antimicrobial agent provides a solution free of bacterial contamination.

Instructions For Use Dilute samples as required before addition to ELISA plate. Highly concentrated samples may need to be diluted 1:100 or more.

References 1. Malinowska, M. *et al.* (2024) Efficacy of a Combination Therapy with Laronidase and Genistein in Treating Mucopolysaccharidosis Type I in a Mouse Model [Int J Mol Sci. 25](#)

[\(4\): 2371.](#)

2. Vanhoutte, L. *et al.* (2017) MRI Assessment of Cardiomyopathy Induced by β 1-Adrenoreceptor Autoantibodies and Protection Through β 3-Adrenoreceptor Overexpression. [Sci Rep. 7: 43951.](#)

Storage	Store at +4°C. DO NOT FREEZE. This product should be stored undiluted.
Guarantee	Guaranteed until date of expiry. Please see product label.
Health And Safety Information	Material Safety Datasheet documentation #10389 available at: https://www.bio-rad-antibodies.com/SDS/OBT1999G 10389
Regulatory	For research purposes only

Related Products

Recommended Useful Reagents

[5x ELISA COATING BUFFER \(BUF030A\)](#)

[10x ELISA WASH BUFFER \(BUF031A\)](#)

[ELISA BSA BLOCK \(BUF032A\)](#)

North & South America	Tel: +1 800 265 7376 Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com	Worldwide	Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com	Europe	Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com
----------------------------------	---	------------------	---	---------------	---

To find a batch/lot specific datasheet for this product, please use our online search tool at: [bio-rad-antibodies.com/datasheets](https://www.bio-rad-antibodies.com/datasheets)
'M350440:190307'

Printed on 23 Feb 2024