

# Datasheet: PUR030 BATCH NUMBER 221112

Description:	PROTEUS MIDI ENDOTOXIN REMOVAL HIGH CAPACITY SPIN COLUMN PACK
Name:	ENDOTOXIN REMOVAL HIGH CAPACITY KIT
Format:	MIDI - Spin Columns
Product Type:	Purification Kit
Quantity:	12 UNITS

### **Product Details**

Appli	ications
-------	----------

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit <a href="www.bio-rad-antibodies.com/protocols">www.bio-rad-antibodies.com/protocols</a>.

	Yes	No	<b>Not Determined</b>	Suggested Dilution
Protein Purification	•			
Where this product has r	not been to	ested for	use in a particular tech	nique this does not

necessarily exclude its use in such procedures.

Preservative	е
Stabilisers	

94% Preservative base of resin containing:

20% Aqueous Ethanol

80% 0.1M Sodium Chloride

## **Endotoxin Level**

This kit can be used for samples with endotoxin loads < 30,000 EU, and is required where up to 1,000,000 EU. For endotoxin loads < 3,000 EU the Mini Endotoxin Removal kit is

recommended.

#### **Product Information**

This kit offers rapid, one step, high grade (>95%) removal of endotoxin from a wide range of protein solutions, with >90% protein recovery.

We recommend reading the <u>Proteus Endotoxin Removal Kit Handbook</u> prior to use.

For more information visit Proteus: Protein Purification Spin Kits.

### Reagents In The Kit

12 x Spin Column Plugs containing Endotoxin Binding Resin

12 x Midi High Capacity Spin Columns (20ml capacity)

12 x 50ml Centrifuge Tubes

Plug Insertion Tool Instructions for use

Additional Reagents Required	0.2µm and 1.2µm Syringe Filter Units Low Endotoxin Pre-equilibration Buffer e.g PBS pH7.4	
Instructions For Use	Instructions for use can be found at www.bio-rad-antibodies.cc	om/uploads/IFU/PUR030.pdf
References	1. Negus, D. <i>et al.</i> (2015) Parenteral Administration of Capsule Prevents Lethal Inhalation Anthrax Infection. <u>Antimicrob Agent AAC.01547-15.</u>	
Storage	Upon receipt of kit, remove resin plugs box and store this at +o DO NOT FREEZE RESIN PLUGS OR STORE AT ROOM TEN Store remainder of kit at room temperature.	
Guarantee	Guaranteed until date of expiry. Please see product label.	
Health And Safety Information	Material Safety Datasheet documentation #qtec44-8 available https://www.bio-rad-antibodies.com/SDS/PUR030 qtec44-8 Not required	at:
Regulatory	For research purposes only	

America

North & South Tel: +1 800 265 7376 Fax: +1 919 878 3751

Email: antibody\_sales\_us@bio-rad.com

Worldwide

Tel: +44 (0)1865 852 700

Europe

Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50

Fax: +44 (0)1865 852 739 Email: antibody\_sales\_uk@bio-rad.com Email: antibody\_sales\_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets 'M350466:190307'

### Printed on 23 May 2025

© 2025 Bio-Rad Laboratories Inc | Legal | Imprint