

Datasheet: PUR030

| Description: | PROTEUS MIDI ENDOTOXIN REMOVAL HIGH CAPACITY SPIN COLUMN PACK |
|----------------------|---|
| Name: | ENDOTOXIN REMOVAL HIGH CAPACITY KIT |
| Format: | MIDI - Spin Columns |
| Product Type: | Purification Kit |
| Quantity: | 12 UNITS |

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

| | Yes | No | Not Determined | Suggested Dilution |
|----------------------|-----|----|----------------|--------------------|
| Protein Purification | | | | |

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures.

| Preservative | |
|--------------|--|
| Stabilisers | |

94% Preservative base of resin containing:

20% Aqueous Ethanol

80% 0.1M Sodium Chloride

Endotoxin Level

This kit can be used for samples with endotoxin loads < 30,000 EU, and is required where up to 1,000,000 EU. For endotoxin loads < 3,000 EU the Mini Endotoxin Removal kit is

recommended.

Product Information

This kit offers rapid, one step, high grade (>95%) removal of endotoxin from a wide range

of protein solutions, with >90% protein recovery.

We recommend reading the Proteus Endotoxin Removal Kit Handbook prior to use.

For more information visit Proteus: Protein Purification Spin Kits.

Reagents In The Kit

12 x Spin Column Plugs containing Endotoxin Binding Resin

12 x Midi High Capacity Spin Columns (20ml capacity)

12 x 50ml Centrifuge Tubes

Plug Insertion Tool Instructions for use

Additional Reagents

0.2µm and 1.2µm Syringe Filter Units

| Required | Low Endotoxin Pre-equilibration Buffer e.g PBS pH7.4 | |
|----------------------------------|--|-----------|
| Instructions For Use | Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/PL | JR030.pdf |
| References | 1. Negus, D. <i>et al.</i> (2015) Parenteral Administration of Capsule Depolymerase E Prevents Lethal Inhalation Anthrax Infection. <u>Antimicrob Agents Chemother. pii: AAC.01547-15.</u> | nvD |
| Storage | Upon receipt of kit, remove resin plugs box and store this at +4°C. DO NOT FREEZE RESIN PLUGS OR STORE AT ROOM TEMPERATURE. Store remainder of kit at room temperature. | |
| Guarantee | Guaranteed until date of expiry. Please see product label. | |
| Health And Safety Information | Material Safety Datasheet documentation #20510 available at: https://www.bio-rad-antibodies.com/SDS/PUR030 20510 | |
| Regulatory | For research purposes only | |

America Fax: +1 919 878 3751

North & South Tel: +1 800 265 7376

Worldwide

Tel: +44 (0)1865 852 700

Europe

Tel: +49 (0) 89 8090 95 21

Fax: +44 (0)1865 852 739

Email: antibody_sales_uk@bio-rad.com

Fax: +49 (0) 89 8090 95 50

Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets 'M433385:241018'

Printed on 18 Oct 2024

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint

Email: antibody_sales_us@bio-rad.com