

Datasheet: PUR032

BATCH NUMBER 21221

| | |
|----------------------|---|
| Description: | PROTEUS MINI ENDOTOXIN REMOVAL SPIN COLUMN PACK |
| Name: | ENDOTOXIN REMOVAL MINI KIT |
| Format: | MINI - Spin Columns |
| Product Type: | Purification Kit |
| Quantity: | 12 UNITS |

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

| | Yes | No | Not Determined | Suggested Dilution |
|----------------------|-----|----|----------------|--------------------|
| Protein Purification | ■ | | | |

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures.

Preservative Stabilisers

94% Preservative base of resin containing:
20% Aqueous Ethanol
80% 0.1M Sodium Chloride

Endotoxin Level

This kit can be used for samples with endotoxin loads < 3,000 EU. For endotoxin loads <30,000 EU both the Midi Standard and High Capacity Endotoxin Removal kits can be used. Endotoxin loads up to 1,000,000 EU require the use of a Midi High Capacity Endotoxin Removal kit.

Product Information

This kit offers rapid, one step, high grade (>95%) removal of endotoxin from a wide range of protein solutions, with >90% protein recovery.

We recommend reading the [Proteus Endotoxin Removal Kit Handbook](#) prior to use.

Reagents In The Kit

12 x vials containing 0.25ml Endotoxin Binding Resin
12 x Mini Spin Column barrels (20ml capacity)
12 x 50ml Centrifuge Tubes
12 x Clear Spin Push Caps
12 x Yellow Incubation Screw Caps
Instructions for use

Additional Reagents

0.2µm syringe filter units

| | |
|--------------------------------------|--|
| Required | Low endotoxin pre-equilibration buffer e.g PBS pH7.4 |
| Instructions For Use | Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/PUR032.pdf |
| References | 1. Negus D <i>et al.</i> (2015) Parenteral Administration of Capsule Depolymerase EnvD Prevents Lethal Inhalation Anthrax Infection. Antimicrob Agents Chemother. pii: AAC.01547-15. |
| Storage | Upon receipt of kit, remove the vials of resin and store them at +4°C. DO NOT FREEZE RESIN VIALS OR STORE AT ROOM TEMPERATURE. Store remainder of kit at room temperature. |
| Guarantee | Guaranteed until date of expiry. Please see product label. |
| Health And Safety Information | Material Safety Datasheet documentation #10508 available at: https://www.bio-rad-antibodies.com/SDS/PUR032 10508 |
| Regulatory | For research purposes only |

| | | | | | |
|----------------------------------|---|------------------|---|---------------|---|
| North & South America | Tel: +1 800 265 7376 Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com | Worldwide | Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com | Europe | Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com |
|----------------------------------|---|------------------|---|---------------|---|

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets
'M350467:190307'

Printed on 23 May 2025