

Datasheet: PUR032 BATCH NUMBER 300514

| Description: | PROTEUS MINI ENDOTOXIN REMOVAL SPIN COLUMN PACK |
|---------------|---|
| Name: | ENDOTOXIN REMOVAL MINI KIT |
| Format: | MINI - Spin Columns |
| Product Type: | Purification Kit |
| Quantity: | 12 UNITS |

Product Details

| App | licatio | ns |
|-----|---------|----|
|-----|---------|----|

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-

rad-antibodies.com/protocols.

| | Yes | No | Not Determined | Suggested Dilution |
|----------------------|-----|----|-----------------------|--------------------|
| Protein Purification | | | | |

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures.

| Preservative | |
|--------------|--|
| Stabilisers | |

94% Preservative base of resin containing:

20% Aqueous Ethanol

80% 0.1M Sodium Chloride

Endotoxin Level

This kit can be used for samples with endotoxin loads < 3,000 EU. For endotoxin loads <30,000 EU both the Midi Standard and High Capacity Endotoxin Removal kits can be used. Endotoxin loads up to 1,000,000 EU require the use of a Midi High Capacity Endotoxin Removal kit.

Product Information

This kit offers rapid, one step, high grade (>95%) removal of endotoxin from a wide range of protein solutions, with >90% protein recovery.

We recommend reading the Proteus Endotoxin Removal Kit Handbook prior to use.

Reagents In The Kit

12 x vials containing 0.25ml Endotoxin Binding Resin

12 x Mini Spin Column barrels (20ml capacity)

12 x 50ml Centrifuge Tubes

12 x Clear Spin Push Caps

12 x Yellow Incubation Screw Caps

Instructions for use

Additional Reagents

0.2µm syringe filter units

| Required | Low endotoxin pre-equilibration buffer e.g PBS pH7.4 |
|----------------------------------|--|
| Instructions For Use | Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/PUR032.pdf |
| References | 1. Negus D <i>et al.</i> (2015) Parenteral Administration of Capsule Depolymerase EnvD Prevents Lethal Inhalation Anthrax Infection. <u>Antimicrob Agents Chemother. pii:</u> <u>AAC.01547-15.</u> |
| Storage | Upon receipt of kit, remove the vials of resin and store them at +4°C. DO NOT FREEZE RESIN VIALS OR STORE AT ROOM TEMPERATURE. Store remainder of kit at room temperature. |
| Guarantee | Guaranteed until date of expiry. Please see product label. |
| Health And Safety Information | Material Safety Datasheet documentation #10508 available at: https://www.bio-rad-antibodies.com/SDS/PUR032 10508 |
| Regulatory | For research purposes only |

North & South Tel: +
America Fax: -

North & South Tel: +1 800 265 7376

Worldwide

Tel: +44 (0)1865 852 700

Europe

Tel: +49 (0) 89 8090 95 21

Fax: +1 919 878 3751
Email: antibody_sales_us@bio-rad.com

Fax: +44 (0)1865 852 739
Email: antibody_sales_uk@bio-rad.com

Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets 'M350467:190307'

Printed on 26 Jun 2024

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint