

Datasheet: RMT1 BATCH NUMBER 166907

Description:	RAT MONOCLONAL ANTIBODY ISOTYPING TEST KIT
Name:	RAT ISOTYPING KIT
Format:	Kit
Product Type:	Kits
Quantity:	10 TESTS

Product Details

Applications	This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit <u>www.bio-rad-antibodies.com/protocols</u> .
	Yes No Not Determined Suggested Dilution Isotyping Assay •
	Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.
Target Species	Rat
Product Information	The rat monoclonal antibody isotyping test kit shows no cross-reactivity with bovine IgG (<0.1 %).
Test Principle	The assay principle is based on mouse polyclonal anti-rat antibodies coupled onto coloured micro particles and equally reactive to any rat monoclonal antibody regardless of its isotype. The isotyping strip has immobilized bands of monoclonal mouse anti-rat antibodies corresponding to each of the common rat antibody isotypes (IgG1, IgG2a, IgG2b, IgG2c and IgM) and to kappa light chains only. One side of the strip bears a positive flow control band, which indicates that the antibody-coated coloured micro particles have migrated through the strip. By using these two components, a rat monoclonal antibody can be screened for isotype by simply diluting the antibody sample, pipetting the diluted sample into the development tube where it forms a complex with the antibody coated micro particles, and inserting the strip. This complex flows through the strip until it is bound by the immobilized mouse anti-rat antibody specific for the rat monoclonal's isotype and its light chain. In approximately 5-10 minutes, the micro particle complex will aggregate as blue bands in the two sections corresponding to the monoclonal antibody's isotype and its light chain. Development of the strip is complete

Reagents In The Kit	1 Desiccant vial containing rat isotyping test strips, 10 tests.
	10 Capped ready-to-use lyophilized microparticle development tubes.

Instructions For Use Note: All reagents should be brought to room temperature before use.

Sample Preparation:

Dilute all monoclonal antibody samples to a concentration of 1.0 ug/ml in PBS containing 1% w/v bovine serum albumin (BSA). If the concentration of the sample is entirely unknown, make dilutions based on the following estimates:

Typically, serum contains between 10-15 mg/ml IgG and ascites can be as high as 10 mg/ml . Hollow fibre bioreactor culture supernatants contain approximately 0.5-1.0 mg/ml, whereas static flask tissue culture supernatants usually contain 10-50 ug/ml. Using these estimates, the appropriate dilutions can be made.

150ul of the diluted sample will be added to the development tubes.

Assay Protocol:

1. Remove the required number of isotyping strips from the desiccant vial and replace thecap. Remove the caps from an equal number of development tubes. **Note:** the tubes may be labeled with a marker for identification.

2. Pipette 150ul of the freshly diluted sample into each development tube and incubate at room temperature for 30 seconds. Vortex the tube briefly to ensure that the coloured micro particle solution is completely re-suspended.

3. Place one isotyping strip, with the solid red end at the bottom, into each development tube.

Interpretation of Results:

Interpret the results at 5-10 minutes once the positive flow control bands have appeared. Within 5-10 minutes, a blue band will appear above the letters in one of the class or subclass windows as well as in the kappa window of the strip, indicating the heavy and light-chain composition of the monoclonal antibody (lambda light chains are very rare in rat monoclonal antibodies, but if present no positive band will be seen in the kappa window). The intensity of the blue bands will increase as the sample continues to flow up the strip. The positive flow control bands on one side of the isotyping test strip should also appear, indicating that the antibody-coated micro particles are functional and have flowed up the strip. In cases where the sample is very dilute, the development time may take up to 10 minutes.

Note: For a permanent experimental record or for an easier interpretation of results when testing multiple samples, the solid red area may be cut off the bottom of the strip to

	prevent further band development once the positive flow control bands have appeared. A gentle stream of air can be applied to the membrane portion of the strip to assist in drying the membrane and preventing any further development. Do not wash the strip to stop the reaction.
References	 Esashi,E. <i>et al</i> (2004) Development of CD4+ macrophages from intrathymic T cell progenitors is induced by thymic epithelial cells. JImmunol. 173: 4360-67 Balyasnikova, I.V. <i>et al</i>. (2005) Monoclonal antibodies to native mouse angiotensinconverting enzyme (CD143): ACE expression quantification, lung endothelial cell targeting and gene delivery. <u>Tissue Antigens. 67: 10-29</u>. Taylor, P.R. <i>et al</i>. (2005) Dectin-2 is predominantly myeloid restricted and exhibits unique activation-dependent expression on maturing inflammatory monocytes elicited <i>in vivo</i>. <u>Eur J Immunol.35: 2163-74</u>. Fukuhara, T. <i>et al</i>. (2006) Functional analysis of nuclear pore complex protein Nup62/p62 using monoclonal antibodies. <u>Hybridoma (Larchmt). 25 (2): 51-9</u>. Koide, A. & Koide, S. (2007) Monobodies: antibody mimics based on the scaffold of the fibronectin type III domain. <u>Methods Mol Biol. 352: 95-109</u>. Hibi, T. <i>et al</i>. (2009) Splicing variations in the ligand-binding domain of ApoER2 results in functional differences in the binding properties to Reelin. <u>Neurosci Res. 63 (4): 251-8</u>. Kato, M. <i>et al</i>. (2010) Preparation of Anti-fragrant Monoclonal Antibodies by the Rat Lymph Node Method and Their Characterization Using Enzyme–linked Immunosorbent Assay <u>J Fac Agr, Kyushu Univ., 55 (1), 91-6</u>. Fichou, N. <i>et al</i>. (2015) Single-Dose Anti-CD138 Radioimmunotherapy: Bismuth-213 is More Efficient than Luteitum-177 for Treatment of Multiple Myeloma in a Preclinical Model. <u>Front Med (Lausanne). 2: 76</u>. Tanaka, M. <i>et al</i>. (2016) A One-Step Immunostaining Method to Visualize Rodent Muscle Fiber Type within a Single Specime. <u>PLoS One. 11 (11): e0166080</u>. Komori, T. <i>et al</i>. (2024) Blockade of OSMRβ signaling ameliorates skin lesions in a mouse model of human atopic dermatitis. <u>FASEB J. 38 (1): e23359</u>.
Storage	Store at +4°C. DO NOT FREEZE. Do not use components from different lots.
Guarantee	Guaranteed until date of expiry. Please see product label.
Health And Safety Information	Material Safety Datasheet documentation #20511 available at: https://www.bio-rad-antibodies.com/SDS/RMT1 20511 Not required
Regulatory	For research purposes only
Technical Advice	PROBLEM

No heavy and or light-chain band appeared on the strip, but the positive flow control bands appeared.

Possible causes:

1. The antibody concentration was too low - prepare a less dilute sample and re-test.

2. No antibody was in the sample - the hybridoma is either not secreting or is not a rat monoclonal. If possible sub-clone the hybridoma and re-test.

3. Freshly diluted samples were not used - prepare fresh dilutions and re-test.

PROBLEM

Multiple heavy and light-chain bands appear on the strip.

Possible causes:

1. Antibody concentration was too high - dilute sample further and re-test.

2. For ascites, there may be small amounts of contaminating antibodies produced - dilute sample further and re-test.

3. For tissue culture supernatant, a mixed culture may be present - re-clone the hybridoma and re-test.

PROBLEM

No positive flow control bands appear.

Possible causes:

1. Sample volume was too low (<150ul) - carefully dilute a fresh sample and pipette 150ul into a new development tube and re-test.

2. Strip removed from development tube too early - re-test and allow strip to react for at least 10 minutes.

North & South Tel: +1 800 265 7376 America Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com

Worldwide Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739

Europe Email: antibody_sales_uk@bio-rad.com

Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com

To find a batch/lot specific datasheet for this product, please use our online search tool at: bio-rad-antibodies.com/datasheets 'M426028:231103'

Printed on 25 Nov 2024

© 2024 Bio-Rad Laboratories Inc | Legal | Imprint