

Datasheet: STAR3000C

Description:	HISTAR DETECTION SYSTEM
Name:	HISTAR DETECTION KIT (HUMAN TISSUES ONLY)
Format:	Kit
Product Type:	Kits
Quantity:	500 TESTS

Product Details

Applications

This product has been reported to work in the following applications. This information is derived from testing within our laboratories, peer-reviewed publications or personal communications from the originators. Please refer to references indicated for further information. For general protocol recommendations, please visit www.bio-rad-antibodies.com/protocols.

	Yes	No	Not Determined	Suggested Dilution
Immunohistology - Frozen	▪			Ready to use
Immunohistology - Paraffin	▪			Ready to use

Where this product has not been tested for use in a particular technique this does not necessarily exclude its use in such procedures. Suggested working dilutions are given as a guide only. It is recommended that the user titrates the product for use in their own system using appropriate negative/positive controls.

Intended Use

Bio-Rad HISTAR Detection System reagents are designed specifically to replace complex immunohistochemical techniques with a simple kit. This kit provides linking and labeling reagents intended for use with species specific primary antibodies for visualizing cellular antigens in tissue specimens. Visualization is accomplished using an ultra-sensitive indirect labeling method that utilizes novel polymer labeling technology. Secondary antibodies are polymerized directly with Horseradish peroxidase (HRP) into compact polymers bearing a high ratio of enzyme to antibody. This biotin-free system offers enhanced sensitivity and minimal background staining.

The HISTAR Detection System is designed for the staining of Human tissue sections. This kit is suitable for use with mouse and rabbit primary antibodies only.

All labeling and blocking reagents are pre-diluted for immediate use, and chromogen is provided in concentrated format. This product is suitable for use with the wide range of Bio-Rad ready-to-use or concentrated primary antibody reagents against human targets. Detailed instructions for use can be found in the complete Bio-Rad PDF datasheet, that can be accessed by clicking on the "Instructions For Use" link. The procedure has been briefly summarized below:

- 1) Prepare tissue sections appropriately according to the requirements of the primary antibody being used e.g. apply antigen retrieval if necessary.
- 2) Incubate paraffin-embedded sections with peroxide blocking reagent for 15 minutes at room temperature. Rinse with Phosphate Buffered Saline (PBS).
- 3) Apply serum block and incubate for 15 minutes.
- 4) Apply primary antibody and incubate for 30 minutes.
- 5) Apply reagent boost and incubate for 20 minutes. Rinse twice with PBS (5 minutes per wash).
- 6) Apply HRP polymer for 30 minutes, then rinse twice with PBS (5 minutes per wash).
- 7) Apply DAB substrate solution for 3-5 minutes, then rinse with distilled water.
[Discard unused substrate solution]
- 8) Counterstain and mount slides using an aqueous mounting medium such as [BUF058A](#).

Reagents In The Kit	VIAL 1, Serum Block 1 x 50ml
	VIAL 2, Reagent Boost 1 x 50ml
	VIAL 3, HRP Polymer 1 x 50ml

Additional Reagents Required	1) Primary antibody and positive control 2) Phosphate Buffered Saline (PBS), pH7.4 (+ 0.2) 3) Hydrogen Peroxide Blocking Reagent 4) Substrate Chromogen (DAB or AEC) 5) Substrate Buffer 6) Counterstain 7) Mounting media
-------------------------------------	--

Bio-Rad can supply a number of these products separately, please see the 'Recommended Useful Reagents' section.

Instructions For Use	Instructions for use can be found at www.bio-rad-antibodies.com/uploads/IFU/STAR3000C.pdf
-----------------------------	---

References	1. Nishimoto, K. <i>et al.</i> (2010) Association of EMCN with Susceptibility to Rheumatoid Arthritis in a Japanese Population. J Rheumatol. 38: 221-8. 2. Dedrick, G.S. <i>et al.</i> (2011) Immunohistochemical study of human costotransverse joints: A preliminary investigation. Clin Anat. 24 (6): 741-7. 3. Alexopoulou, Z. <i>et al.</i> (2016) Deubiquitinase Usp8 regulates α -synuclein clearance and modifies its toxicity in Lewy body disease. Proc Natl Acad Sci U S A. Jul 21. pii: 201523597. [Epub ahead of print]
-------------------	---

Storage	Store at +4°C. DO NOT FREEZE. This product should be stored undiluted.
----------------	---

Guarantee Guaranteed until date of expiry. Please see product label.

Health And Safety Information Material Safety Datasheet documentation #10362 #10282 #10456 available at:
Serum Block (10362): <https://www.bio-rad-antibodies.com/uploads/MSDS/10362.pdf>
Reagent Boost (10282): <https://www.bio-rad-antibodies.com/uploads/MSDS/10282.pdf>
HRP Polymer (10456): <https://www.bio-rad-antibodies.com/uploads/MSDS/10456.pdf>

Regulatory For research purposes only

Related Products

Recommended Useful Reagents

[PEROXIDE BLOCKING REAGENT \(BUF017B\)](#)
[10X PHOSPHATE BUFFERED SALINE \(BUF036A\)](#)
[PERMANENT AQUEOUS MOUNTING MEDIUM \(BUF058C\)](#)

North & South America	Tel: +1 800 265 7376 Fax: +1 919 878 3751 Email: antibody_sales_us@bio-rad.com	Worldwide	Tel: +44 (0)1865 852 700 Fax: +44 (0)1865 852 739 Email: antibody_sales_uk@bio-rad.com	Europe	Tel: +49 (0) 89 8090 95 21 Fax: +49 (0) 89 8090 95 50 Email: antibody_sales_de@bio-rad.com
----------------------------------	---	------------------	---	---------------	---

From March 15, 2021, we will no longer supply printed datasheets with our products.
Look out for updates on how to access your digital version at [bio-rad-antibodies.com](https://www.bio-rad-antibodies.com)

'M352428:190327'

Printed on 10 Feb 2021

© 2021 Bio-Rad Laboratories Inc | [Legal](#) | [Imprint](#)